

# Urisys 1100<sup>®</sup> Urine Analyzer

cobas<sup>®</sup>

## Policies and Procedures

for use with Urisys 1100<sup>®</sup> Urine Analyzer

**This is a CLIA-Waived System**

Roche Diagnostics is not responsible for any changes made to this diskette by the customer.



## Introduction

Roche Diagnostics is pleased to assist you in developing policies and procedures and implementing regulations of the Clinical Laboratory Improvement Amendment of 1988 (CLIA '88) in your facility. This guide contains sample policies and procedures for use with the Urisys 1100® Urine Analyzer. These sample policies and procedures are intended only as a guide, and it is important that you make whatever adjustments may be required as a result of a change in policy or a change in law or regulations or their interpretation.

A total quality management program should include written policies and testing procedures that are customized for your particular facility. Any policies and procedures that involve using Chemstrip urine test strips should be written prior to the start of any urine testing.

This guide provides sample policies and procedures that meet the Section 493.1211 requirements. Blank spaces let you customize the policies and procedures with information specific to your facility. You are encouraged to highlight important areas and procedures currently in use.

We suggest you store a *Urisys 1100® Urine Analyzer Operator's Manual*, *Chemstrip urine test strip package insert* and *control package insert* in the *Product Information* section for quick reference. You may also want to include the following clinical guidelines and regulations in the *Regulatory Requirements* section.

- JCAHO Guidelines
- CAP Guidelines
- CLIA '88 Guidelines
- State Guidelines
- Bloodborne Pathogen Standard

If you have any questions, please call Roche Diagnostics Technical Service Center at 1-800-428-4674, available 7 days per week, 24 hours a day, and 365 days per year.



## **Purpose**

To establish safe, accurate urine test results to be used for screening and monitoring patient status.

## **General Policies and Procedures**

- The policies and procedures pertaining to urinalysis with the Urisys 1100® Urine Analyzer are reviewed at least once a year.
- Any changes to the text as they apply to methodology or testing policy are signed by the appropriate authority, dated, and inserted in the policies and procedures manual.
- Obsolete or superseded procedures are replaced in the manual, but are retained for compliance purposes.
- The most recent package inserts and the *Urisys 1100® Urine Analyzer Operator's Manual* provided by Roche Diagnostics should be referenced for the most recent procedural information.

# Table of Contents

- I. Policies and Procedures..... 1
  - A. Principle of Operation..... 1
  - B. Specimen Collection..... 3
  - C. Test Strip Storage and Stability ..... 4
  - D. Calibration of the Urisys 1100® Urine Analyzer..... 5
  - E. Testing with the Urisys 1100® Urine Analyzer..... 7
  - F. Documentation of Urinalysis Results..... 11
  - G. Expected Values..... 12
  - H. Quality Control Procedure..... 13
  - I. Documentation of Quality Control Results ..... 15
  - J. Limitations of the Procedure ..... 16
- II. Preventive Maintenance ..... 18
- III. Infection Control Guidelines..... 20
- IV. Operator Certification/Recertification ..... 21
- V. Appendix..... 25
- VI. Log Sheets ..... 28
- VII. Regulatory Requirement..... 44
- VIII. Product Information..... 45



# I. Policies and Procedures

## A. Principle of Operation

The Roche Diagnostics Urisys 1100® Urine Analyzer is a semi-automated analyzer intended for *in vitro* semi-quantitative determination of urine analytes. The analyzer is designed to read and evaluate the results of Chemstrip 10MD\*, Chemstrip 7, and Chemstrip 5 OB Urine Test Strips. These urine test strips are multiparameter strips used for the determination of specific gravity, pH, leukocytes, nitrite, protein, glucose, ketones, urobilinogen, bilirubin and blood in urine.

The Urisys 1100® Urine Analyzer is a reflectance photometer. It reads the urine test strips under standardized conditions, saves the results to memory and outputs them via its own built-in printer and/or serial interface.

The Urisys 1100® Urine Analyzer standardizes urine test strip results by eliminating factors known to affect visual evaluation of urine test strips, such as; variable lighting conditions at the workplace, individual skill levels at matching test strip pad colors, failure to keep the prescribed testing times, and clerical errors.

The test strip is placed on a sliding tray, and a stepping motor moves it under the reading head, which remains stationary. The analyzer reads the reference pad, followed by each of the test pads on the strip. The reading head contains LEDs that emit light at various wavelengths. Reading is done electro-optically as follows:

The LED emits light of a defined wavelength onto the surface of the test pad at an optimum angle. The light hitting the test zone is reflected proportionally to the color produced on the test pad and is picked up by the detector, a phototransistor positioned directly above the test zone. The phototransistor sends an analog electrical signal to an A/D converter, which changes it to digital form. The microprocessor then converts this digital reading to a relative reflectance value by referring it to a calibration standard.

Finally, the system compares the reflectance value with the defined range limits (reflectance values which are programmed into the analyzer for each parameter) and outputs a semi-quantitative result.

Each test pad is read photometrically after about 55-65 seconds. In strongly alkaline urine samples, the Urisys 1100® Urine Analyzer automatically corrects the result of the specific gravity test.

The following is a list of test principles specific to each parameter. Please reference the most recent *Chemstrip urine test strip package insert* for modifications in test principles.

- **Specific Gravity:** In the presence of cations, protons are released by a complexing agent in the test pad. The indicator bromthymol blue changes from blue via blue-green to yellow.
- **pH Test:** The test pad contains the indicators methyl red and bromthymol blue. These indicators give clearly distinguishable colors over the pH range of 5 to 9.<sup>1,2</sup> Colors range from orange through yellow and green to blue.
- **Leukocyte Test:** Granulocytic leukocytes contain esterases which catalyze the hydrolysis of an indoxylcarbonic acid ester to indoxyl. The indoxyl formed reacts with a diazonium salt to produce a purple color.
- **Nitrite Test:** Nitrite, if present, reacts with an aromatic amine to give a diazonium salt, which, by coupling with a further compound, yields a red-violet azo dye.<sup>3,4,5</sup>
- **Protein Test:** The test is based on the color change of the indicator 3', 3'', 5', 5''-tetrachlorophenol-3, 4, 5, 6-tetrabromosulfophthalein in the presence of protein. A positive reaction is indicated by a color change from yellow to light green/green.<sup>6,7,8</sup>

\*Hospitals may use Chemstrip 10UA Urine Test Strips.

- **Glucose Test:** Glucose detection is based on the enzymatic glucose oxidase/oxidase (GOD/POD) method. The reaction utilizes the enzyme glucose oxidase to catalyze the formation of gluconic acid and hydrogen peroxide from the oxidation of glucose. In turn, a second enzyme, peroxidase, catalyzes the reaction of hydrogen peroxide with the chromogen tetramethylbenzidine to form a green dye complex. A positive reaction is indicated by a color change from yellow to green.<sup>9, 10</sup>
- **Ketone Test:** Based on the principle of Legal's Test, sodium nitroprusside and glycine react with acetoacetate and acetone in an alkaline medium to form a violet dye complex. A positive result is indicated by a color change from beige to violet.<sup>11, 12</sup>
- **Urobilinogen Test:** Urobilinogen is coupled with 4-methoxybenzene-diazonium-tetrafluoroborate in an acid medium to form a red azo dye.<sup>13</sup>
- **Bilirubin Test:** The detection of bilirubin is based on the coupling reaction of a diazonium salt with bilirubin in an acid medium. The reaction yields a pink to red-violet color proportional to the total bilirubin concentration.<sup>14</sup> (Some users may describe this as a cream to peach color.)
- **Blood Test:** Hemoglobin and myoglobin, if present, catalyze the oxidation of the indicator by the organic peroxide contained in the test pad. Intact erythrocytes hemolyze on the test pad and liberate hemoglobin, which produces a green dot. Since the test pad absorbs several microliters of urine, more erythrocytes become visible than would correspond to 1 $\mu$ L.<sup>15-19</sup> Separate sets of color blocks are representative of erythrocytes and hemoglobin. Scattered or compacted green dots on the yellow test pad are indicative of intact erythrocytes, or myoglobin.
- **Compensation Pad:** The compensation area is not impregnated with reagent. This pad allows the instrument to compensate for the intrinsic color of urine, which may affect test results.

## **B. Specimen Collection**

### **Policy**

- Use a freshly voided urine specimen, first-morning urine specimen, or post-prandial urine specimen.
- When testing urine specimens, every precaution must be taken to avoid contamination of the specimen and the operator.
- Because of the potentially hazardous nature of handling urine and the possibility of its contamination with blood, it is recommended that disposable latex gloves and laboratory coats be used when collecting specimens. Consult your facility's infection control policy.
- Instructions for collection of a clean-catch midstream urine specimen may be found in *Section V. Appendix*.

### **Procedure**

1. Follow physician's order(s).
2. Assemble appropriate supplies.
  - Cleansing towelette
  - Water-saturated sponge
  - Clean urine container with lid
3. Instruct patient to wash genital area (do not use towelettes with strong cleansing agents), rinse with water, and then obtain a clean-catch midstream urine sample in the clean container. Refer to *Section V. Appendix* for instructions.
4. Test urine as soon as possible after collection to avoid deterioration of specimen. If specimen cannot be tested within *two* hours of collection, refrigerate immediately (at 2°-8°C) in a closed container. Bring specimen to room temperature before testing. Mix thoroughly before testing.
5. If urine is not tested immediately, mix specimen thoroughly by capping container and swirling several times before testing to avoid incorrect results. Do not centrifuge or add preservatives to the urine specimen.

## **C. Test Strip Storage and Stability**

### **Policy**

- Store Chemstrip urine test strips at 2 °C (36 °F) – 30 °C (86 °F). Do not freeze.
- Opened Chemstrip urine test strips are stable until the expiration date on the vial label when stored in the original capped vial. The vial must be closed immediately after use, using the original cap.

## D. Calibration of the Urisys 1100<sup>®</sup> Urine Analyzer

(with Chemstrip 10MD\* Test Strips)

### Policy

- Calibration is required if using the Chemstrip 10MD\* Test Strips. (When using Chemstrip 5 OB or Chemstrip 7 Test Strips, the calibration procedure is not required. These test strips are calibrated using the gray reference pad on the test strip tray.)
- The Urisys 1100<sup>®</sup> Urine Analyzer must be calibrated every seven days or when indicated by the analyzer.
- The Chemstrip Urine Analyzer Calibration Strips (Catalog Number 418007) should remain in the vial until just before use. Do not touch the elevated gray areas on the strips.
- The calibration strips should be used only *once*.
- If you do not obtain a successful calibration, the analyzer will not process samples.
- Calibration should be documented on the *Preventive Maintenance Log*.

### Procedure

1. Plug the power cord provided with the analyzer into the power outlet at the rear of the analyzer and then into a wall outlet. Turn the analyzer on. The analyzer will then automatically perform a self check. If the self check is completed successfully, the message “self check OK” prints along with the time and date.

2. Make sure the test strip tray is clean and dry.

Note: After a successful self check, the operator may have to enter an operator ID and password (if these options have been selected).

3. If the message “REPEAT CALIBRATION” is in the display, press START. If the analyzer is in Ready-to-Measure mode, press the left function key to select “Calibr”. The message, “START CALIBRATION” appears on the display.

4. Remove a calibration strip from its container, being careful not to touch the pads. Do not allow the calibration strip to come into contact with urine.

5. With the retaining bar open, place the calibration strip, pads facing upward, onto the test strip tray and insert the front edge of the strip under the plastic clip.

6. Press the START button. You will hear a beep and then the tray advances slightly, the retaining bar closes, and the calibration strip is read. When calibration is complete, the tray moves to the start position and the retaining bar opens.

7. Remove the calibration strip and discard. Remember that each calibration strip is used only once.

8. If the calibration is valid, results are stored in memory, and a report with the reflectance values, time, and date automatically prints. Reflectance values for positions 1 through 11 for the orange LED are printed in the middle column and for the green LED in the right column.

\*Hospitals may use Chemstrip 10UA Urine Test Strips.

9. If the results for the reference pad or the calibration strip are outside the programmed tolerances, one of the following messages appears on the display:
  - “REFERENCE PAD ERROR”
  - “CALIBRATION INVALID” or
  - “CALIBRATION ERROR”

If a calibration error occurs, repeat the calibration procedure using a new calibration strip.

1. Press the START button and the display will return to the “START CALIBRATION” menu.
2. Place the new calibration strip on the test strip tray and press the START button again.
3. If the calibration is valid, the message “CALIBRATION OK” prints.
4. If the calibration fails a second time and a calibration error appears on the display, refer to the Operator’s Manual or call Roche Diagnostics Technical Service at 1-800-428-4674, 7 days a week, 24 hours a day, and 365 days per year.

## E. Testing with the Urisys 1100<sup>®</sup> Urine Analyzer

### Policy

- Only a certified operator may perform a urine test with the Urisys 1100<sup>®</sup> Urine Analyzer.
- Because of the potentially hazardous nature of handling urine and the possibility of its contamination with blood, it is recommended that disposable latex gloves and laboratory coats be used when collecting specimens and performing test procedures. Consult your facility's infection control policy.
- Urine should be handled at Biosafety Level 2 as recommended for any potentially infectious material in the Centers for Disease Control/National Institutes of Health manual, *Biosafety in Microbiological and Biomedical Laboratories*, 1988.<sup>24</sup> Universal precautions may apply if the urine is contaminated with blood or if required by the infection control policy of the facility.
- Outdated Chemstrip urine test strips are discarded.

### Procedure

1. The following equipment is assembled prior to testing:

Urisys 1100<sup>®</sup> Urine Analyzer

Chemstrip urine test strips

Patient specimen

Disposable latex gloves (if required by facility)

Laboratory coat (if required by facility)

Absorbent paper

2. Check the expiration date on the Chemstrip urine test strip vial prior to testing.
3. Plug the power cord provided with the analyzer into the power outlet at the rear of the analyzer and then into a wall outlet. Turn the analyzer on. The analyzer will then automatically perform a self check. If the self check is completed successfully, the message "self check OK" prints along with the time and date.

Note: After successful self check, the operator may have to enter their operator ID and password if these options have been activated.

4. The Urisys 1100<sup>®</sup> Urine Analyzer is ready for measurement as long as valid calibration values are stored in the analyzer. Refer to *Section I.D. Calibration of the Urisys 1100<sup>®</sup> Urine Analyzer* of this manual.
5. Verify the test strip tray is clean of any residue. If there is residue on the tray, refer to *Section II. Preventive Maintenance* of this manual for the cleaning procedure. Urine residue on the test strip tray may result in inaccurate results.
6. If the sample has been refrigerated, allow it to come to room temperature. Mix the specimen by capping the container and swirling several times if not freshly voided.

7. Remove the cap from the test strip vial and remove one Chemstrip urine test strip.
8. Dip the test strip for one second in the urine sample. Draw the edge of the test strip along the rim of the specimen container. Gently touch (one second) the long edge of the test strip to a piece of absorbent paper to remove any excess urine.
9. If the analyzer is in Standby mode, press START to return to Ready-to-Measure mode. The test strip tray and the retaining bar must be in the open position. Place the test strip, with the pads facing upward, onto the test strip tray and insert the front edge of the strip under the plastic clip. It is important that the strip is correctly positioned and the START button is pressed within 5-10 seconds of dipping the strip.
10. Press the START button. You will hear a beep and then the tray advances slightly, and the retaining bar closes. First the gray reference pad located on the test strip tray is read.
11. If the test strip is not lined up straight on the test strip tray, gently move it until it is properly lined up under the retaining bar. Be careful not to move the tray.
12. Approximately 60 seconds after the START key is pressed, all test strip pads will be read. The results will print and the next sample number appears on the display.
13. The test strip tray will return to the start position and the retaining bar will open. Remove the used test strip and wipe the test strip tray as needed with a lint-free cloth to remove any urine residue. Refer to *Section II. Preventive Maintenance* of this manual for the cleaning procedure.
14. Properly dispose of the urine, urine container and test strip.

## Guidelines

- Mechanical problems are evident when the analyzer displays an error message. A chemistry problem may become evident with an unexpected result. Abnormal results are flagged with an asterisk. If you obtain an unexpected result on any chemistry, refer to the *Limitations and Performance Characteristics* sections of the *Chemstrip urine test strip package insert*.
- If you are comparing the analyzer values against the visual strip values and the deviation is more than one concentration block, one of the following may have occurred:
  - Test strip is not positioned correctly in the test strip tray
  - Power supply failure or high frequency interference by other devices
  - Visual strip is deteriorated
  - Improper testing technique

If one of the above situations has occurred, repeat the measurement with a new test strip, being very careful to properly position the test strip. Remove any external source of interferences.

- The following is a list of error messages and suggestions on how to eliminate the problem:

<b>Error Code</b>	<b>Error Description</b>	<b>Cause</b>	<b>Action</b>
E1	Reference pad error	Gray reference pad on the test strip tray soiled or damaged.	Switch off the instrument. If not visibly damaged, clean and dry the pad. Insert the tray and wait for the self-check to finish. If the error message reoccurs, call Roche Diagnostics Technical Service Center.
E2	Wrong strip	The test strip used is different from the one for which the analyzer has been programmed for (either Chem 10, Chem 5, or Chem 7). See Operator's Manual.	Press the START button. Repeat the measurement using the test strip that the instrument has been programmed for.
E3	Strip measurement	a) No test strip present on the tray. b) Test strip is incorrectly positioned on the tray. c) Urine on the test strip has dried. d) Test strip has not been dipped in urine.	Press the START button. Repeat the measurement with a new test strip. Make sure the strip is inserted correctly and that the retaining bar properly closes after the START button is pressed.
E4	Calibration Error	Calibration values differ from those obtained in the last valid calibration.	Press the START button. Repeat the calibration using a new calibration strip. Make sure the calibration strip is properly positioned under the clip on the test strip tray.
E5	Calibration Invalid	Calibration values out of tolerance.	Check the reference pad for soil or damage. Clean the reference pad or use the spare tray. Repeat the calibration with a new calibration test strip. If the error reappears, call Roche Diagnostics Technical Service Center.
E6	Chip Error	The chip module underneath the printer cover is missing, not making contact, or is defective.	Switch off the analyzer, insert the chip module and switch the instrument on again. If the error message appears again, call Roche Diagnostics Technical Service Center.
E7	Missing Tray	Test strip tray is missing or not inserted far enough to engage the motor.	Insert the test strip tray correctly. (See Section 4 of Operator's Manual.) Press the START button.
E8	Tray Position Error	The positioning hole in the tray is dirty or still wet after cleaning; the retaining bar is open while the tray is advancing; or the retaining bar mechanism is contaminated with urine and blocked.	Clean, blow through, or dry the positioning hole on the test strip tray using a lint-free cloth. Remove any urinary deposits on both the top and bottom of the test strip tray. Insert the tray again and press the START button. If the message error persists, call Roche Diagnostics Technical Service Center.
E9	Wrong tray	The test strip tray is not the one programmed for the analyzer or the gray reference pad is missing from the tray. Gray reference pad is scratched or dirty.	Press the START button. The strip type menu is displayed. The strip type must match the tray type. (See Operator's Manual, Section 4.2.) Use the correct tray and make sure the gray reference pad is clean. Press the START button.
E10	Light Barrier Error	The light barrier used to control the position of the test strip is defective or the tray transport is blocked.	Pull out the tray and return it to the start position. Press the START button. If the error message persists, call Roche Diagnostics Technical Service Center.
E11	Motor Step Error	The stepping motor drive is out of tolerance or the advance of the test strip tray is blocked.	Carefully clean the tray. Clean and dry the positioning hole using a lint-free cloth. Remove any urinary deposits on both the top and bottom of the test strip tray. Insert the tray again and press the START button. If the message error persists, call Roche Diagnostics Technical Service Center.
E12	Optics Error	The reference pad is missing from the test strip tray or an LED or phototransistor is defective.	Attach the reference pad. Press the START button. If the error message appears again, call Roche Diagnostics Technical Service Center.
E13	Printer Error	Printer connection fault or paper jam. This message will not appear if there is no paper.	Switch off the analyzer. If there is a paper jam, clear it. Switch the analyzer on again. If the error message appears again, call Roche Diagnostics Technical Service Center.

<b>Error Code</b>	<b>Error Description</b>	<b>Cause</b>	<b>Action</b>
E14	Interface Error	Unable to transfer data to PC or host in bi-directional mode.	Check the data cable. Verify that the PC or host is ready to receive data. Use the "Send" function to transfer data or press "Home" to resume Ready-to-Measure status.
E15 REFERENCE PAD ERROR BOTTOM!		The bottom portion of the reference pad on the tray is soiled or damaged.	see E1
E16 REFERENCE PAD ERROR TOP!		The top portion of the reference pad on the tray is soiled or damaged.	see E1
E17 INVALID PASSWORD!		The entered password doesn't match.	Enter an correct password.
E18 INVALID OP.ID!		The entered Operator ID is not valid.	Enter a valid Operator ID.
E19 LIST DOWNLOAD FAILED		The new Operator ID list download failed.	No action. After 2 seconds next state starts with old list if there was.
E20 NO VALID LIST!		There is not a valid list at all in device.	Try to download an Operator ID list from the host or continue without authenticated operator.
Check Measurement	Analyzer prints out software and chip version number and 3-digit numbers without naming the parameters	Service function is activated.	Press the "Back" function key to return to the main menu.
	Values obtained do not compare with those from visual examination.	a) Test strip incorrectly positioned on the test strip tray. b) Uncharacteristic test pad colors. c) The wrong test strip, such as the calibration strip, was used. d) Electromagnetic interference from other devices. e) Improper storage or technique.	Repeat the measurement with a new test strip. Follow the procedure carefully and make sure the test strip is inserted properly. Repeat calibration, if necessary. Move the analyzer to another location if external interference is suspected.
	No printout	"Printer: Off" is selected or the printer/software is defective.	Choose "Printer: On" to reactivate the printer. Request a patient report through the "Print" function. If this fails, activate the "Linefeed" function. If there is still no response, call ROCHE DIAGNOSTICS Technical Service Center.
	The analyzer will not read, even though the sequence number is displayed.		Switch the analyzer off and back on again.

## F. Documentation of Urinalysis Results

### Policy

- The date, time, initials of the operator, patient name/patient ID number, and parameter values are recorded on the patient chart or the appropriate log for your facility.
- Test requisitions, test authorizations, and test results are retained for a minimum of two years.
- Test requisition or authorization includes:
  - Patient's name
  - Name and address (or other identifier) of the person requesting the test
  - Test to be performed
  - Date the sample was obtained
  - Date the test was performed
- An audit trail exists, linking the patient test with the Chemstrip test strip lot number and control lot number used.

### Procedure

1. Record the date, time, initials of the tester, patient name/patient ID number, and test values on the *Urinalysis Report Form* of this guide or the appropriate log for your facility.
2. Indicate the type of specimen (first morning, random, clean-catch midstream, etc.) on the appropriate log.

**(See Section VI. Log Sheets for sample log sheets.)**

## G. Expected Values

### Guidelines

Please refer to the most recent *Chemstrip urine test strip package insert* for any changes to the expected values information.

- **Specific Gravity:** Random urine specimens vary from 1.001 to 1.035. Twenty-four hour urine specimens from normal adults with normal diets and fluid intake will have a specific gravity of 1.016 to 1.022.<sup>22</sup>
- **pH:** Urine pH values generally range from 5 to 9. The most frequent pH values for first-morning specimens in healthy subjects are between 5 and 6.
- **Leukocytes:** Normal urine should produce no color reaction. “Trace” indicates a possible borderline situation, and the test should be repeated on a fresh urine sample from the patient. Positive and repeated trace findings indicate the need for further testing of the patient and/or urine sample in accordance with the medically accepted procedures for pyuria.
- **Nitrite:** A nitrite concentration as low as 0.05 mg/dL will produce a slightly pink coloration of the test pad. This indicates a positive result.
- **Protein:** A color change from yellow to light green/green will occur if protein is present in urine. The concentrations given on the vial label correspond with the albumin concentration in urine. Pathological proteinuria usually will produce persistent values above 30 mg/dL. Clinical significance of the trace result should be determined by additional testing.
- **Glucose:** Due to the test’s sensitivity, glucose should not be detectable in normal urine. Therefore, any positive reaction should be followed by further diagnostic evaluation of the patient, such as a quantitative blood glucose or a glucose tolerance test.
- **Ketones:** Ketone bodies should not be detected in normal urine with this test. Fasting or starvation diets may cause positive results. In known pathological conditions such as diabetes, the presence of ketones may be useful as an index of metabolic status.
- **Urobilinogen:** Concentrations are usually greater in the afternoon than during the remainder of the day. Values up to 1 mg/dL are usually considered normal.<sup>13</sup>
- **Bilirubin:** In normal urine, bilirubin should not be detectable with this test. However, the test is very sensitive to bilirubin (0.5 mg/dL will produce positive results), and any positive reaction indicates that further diagnostic evaluation of the patient is needed.
- **Blood:** Erythrocyte excretion up to 5 Ery/ $\mu$ L may be expected in normal urine.<sup>17,18</sup> Levels above this may require further evaluation of the patient.

## H. Quality Control Procedure

### Policy

- Quality control testing will be performed with a minimum of two (2) quality control solutions according to your facility's policy.
- It is required to run a positive and negative control daily and when a new vial of test strips is opened (including every lot change). Values obtained for these controls should fall within the limits established by the laboratory or the control manufacturer.
- Check the appropriate boxes for your facility's additional quality control frequency, being sure to meet all standards of your regulatory agencies.
  - Each \_\_\_\_\_ (when? day, shift, etc.).
  - When the test strips have been exposed to extreme heat, humidity or cold.
  - When test results contradict clinical symptoms.
  - As a way of checking operator competency.
  - Other \_\_\_\_\_.
- If a quality control test result falls within the acceptable control range, it is acceptable to proceed with patient testing.
- If a quality control test result falls outside of the acceptable control range, repeat the test with a fresh Chemstrip urine test strip. Check the stability of the control being used. The problem must be corrected before proceeding with patient testing.
- Any quality control result that falls outside of the acceptable control range, along with any corrective action to restore that result to acceptable range, must be recorded in the *Urinalysis Quality Control Log* or other quality control log.
- The controls should be stored at the proper temperatures according to the manufacturer's recommendations.
- Any outdated test strips or controls will be discarded.
- The urine test strip lot number and expiration date are found on the label of each vial of Chemstrip urine test strips. This information should be recorded in the *Urinalysis Quality Control Log* or other quality control log.
- Chemstrip urine test strips must be stored at temperatures below 30°C (86°F). Do not freeze. Test strips are stable until the expiration date. Test strips must be stored in the same capped vial in which they were packaged, and the vial cap must be immediately replaced after removal of a test strip.
- Urine controls recommended for use with the Urisys 1100® Urine Analyzer are KOVA-Trol® and KOVA® Liqua-Trol™ (HYCOR Biomedical Inc.), the Dipper Urine Dipstick Control (Quantimetrix) and Liquichek™ (BIORAD). Refer to *Section V. Appendix* for more information about these controls.
- Insert your facility's quality control preparation information after this page.

## Procedure

1. The following equipment is needed for quality control testing:

Urisys 1100® Urine Analyzer

Chemstrip urine test strips

Commercially prepared urinalysis control

Disposable latex gloves (if required by facility)

Laboratory coat (if required by facility)

Absorbent paper

2. Prepare controls according to the manufacturer's instructions.
3. Check the expiration date on the Chemstrip urine test strip vial prior to testing.
4. The Urisys 1100® Urine Analyzer is ready for measurement as long as valid calibration values are stored in the analyzer. Refer to *Section I.D. Calibration of the Urisys 1100® Urine Analyzer* of this manual.
5. Verify the test strip tray is clean of any residue. If there is residue on the tray, refer to *Section II. Preventive Maintenance* of this manual for the cleaning procedure. Any residue on the test strip tray may result in inaccurate results.
6. Remove the cap from the test strip vial and remove one Chemstrip urine test strip.
7. Dip the test strip for one second in the control. Draw the edge of the test strip along the rim of the container. Gently touch (one second) the long edge of the test strip to a piece of absorbent paper to remove any excess control.
8. If the analyzer is in Standby mode, press START to return to Ready-to-Measure mode. The test strip tray and the retaining bar must be in the open position. Place the test strip, with the pads facing upward, onto the test strip tray and insert the front edge of the strip under the plastic clip.
9. Press the START button. You will hear a beep and then the tray advances slightly and the retaining bar closes. First the gray reference pad located on the test strip tray is read.
10. If the test strip is not lined up straight on the test strip tray, gently move it until it is properly lined up under the retaining bar. Be careful not to move the tray.
11. Approximately 60 seconds after the START key is pressed, all test strip pads will be read. The results will print and the next sample number appears on the display.
12. The test strip tray will return to the start position and the retaining bar will open. Remove the used test strip and wipe the test strip tray as needed with a lint-free cloth to remove any residue. Refer to *Section II. Preventive Maintenance* of this manual for the cleaning procedure.
13. Properly dispose of the control solution and test strip.

# I. Documentation of Quality Control Results

## Policy

- Quality control records are retained for a minimum of two years.
- The date, time, initials of the operator, and quality control result are recorded in the *Urinalysis Quality Control Log* or other quality control log. The log should also include the lot number and expiration date of the control solutions and urine test strips.
- Any quality control result that falls outside the acceptable control range, along with any corrective action to restore that result to acceptable range, is recorded in the *Urinalysis Quality Control Log* or other quality control log.
- An appropriate authority or an appointed individual reviews the *Urinalysis Quality Control Log* for completeness and notes any trends that could indicate potential problems. Such trends include gradual drifting of values, sudden shifts in urine control values while using the same lot of strips, and operator performance. The log is reviewed \_\_\_\_\_ (when? weekly, monthly, quarterly).
- An audit trail should exist to link the control test values to the patient test values performed that day by listing the appropriate Urisys 1100<sup>®</sup> Urine Analyzer, Chemstrip urine test strip and control solution lot number used.
- Insert your facility's policy for who is responsible for the performance and documentation of quality control.

---

---

---

---

---

---

---

---

## J. Limitations of the Procedure

Please reference the most recent *Chemstrip urine test strip package insert* for modifications in the limitations of procedure.

- **Specific Gravity:** The chemical principle of the test may cause slightly different results compared with other specific gravity methods when elevated amounts of certain constituents are present. Glucose and urea concentrations greater than 1% may cause a low specific gravity reading relative to other methods. In the presence of moderate amounts of protein (100-500 mg/dL) or ketoacidosis, readings may be elevated.
- **Leukocyte:** This test is not affected by erythrocytes in concentrations up to 10,000/ $\mu$ L or by bacteria common in urine. Do not collect specimens in containers that have been cleaned with strong oxidizing agents. Do not use preservatives. The drugs cephalexin and gentamicin have been found to interfere with this test. Nitrofurantoin colors the urine and thus causes interference with visual interpretation of the test strip. High levels of albumin ( $>500$  mg/dL) in the urine may interfere with the test results.
- **Nitrite:** Large amounts of ascorbic acid (see under glucose) decrease the sensitivity of the test. False positive readings may be produced by medication that colors the urine red or that turns red in an acid medium (e.g., phenazopyridine).
- **Protein:** False positive results may be found:
  - In strongly alkaline urine (pH of 9 or higher).
  - During therapy with phenazopyridine.
  - When infusions of polyvinylpyrrolidone (blood substitutes) are administered.
  - When residues of disinfectants containing quaternary ammonium groups or chlorohexidine are present in the urine container.
- **Glucose:** The effect of ascorbic acid (vitamin C) retained in the urine due to ingestion of vitamin tablets, antibiotics, or fruit juices has been eliminated at glucose concentrations of 100 mg/dL and above so that false negative readings rarely occur, even at high concentrations of ascorbic acid. False positive readings may be produced by strongly oxidizing cleaning agents in the urine container.
- **Ketone:** Red-orange to red color shades, which are readily distinguished from the colors obtained with ketone bodies, can be produced by phenylketone or phthalein compounds administered for liver and kidney function tests. 2-Mercaptoethane sulphonate sodium (MESNA) or other sulfhydryl-containing compounds may cause false positive results.<sup>21</sup>
- **Urobilinogen:** The total absence of urobilinogen cannot be detected. Most urine specimens give a slight pink reaction. This test gives the same color reaction with urobilinogen as with stercobilinogen; however, the differentiation is not of diagnostic importance. Urine from patients treated with phenazopyridine may show a false positive reaction. Nitrite concentrations above 5 mg/dL or formalin concentrations above 200 mg/dL (used as a preservative) may cause a decrease in color reaction.

- **Bilirubin:** Large amounts of ascorbic acid present in the urine following the ingestion of medication containing vitamin C or fruit juices lower the sensitivity of the test. In case of doubt, repeat test on urine voided at least 10 hours after the last administration of vitamin C. Elevated concentrations of nitrite, as in urinary tract infections, may result in lower bilirubin values. Large amounts of urobilinogen in the urine affect the color change of the bilirubin test, but not enough to give a positive result. False positive results may be produced by medications that color the urine red or turn red in an acid medium (e.g., phenazopyridine).
- **Blood:** False negative readings are obtained when formalin is used to preserve the urine. Nitrite in excess of 10 mg/dL in the urine (which is rare in urinary tract infections) delays the reaction. False positive results can be produced by residues of strongly oxidizing cleaning agents in the urine container. Urine from menstruating females occasionally yields a positive result. This test has not been found to be affected by the ingestion of reasonable quantities of ascorbic acid.

## II. Preventive Maintenance

### Policy – Cleaning the Analyzer

- Clean and disinfect the exterior of the Urisys 1100<sup>®</sup> Urine Analyzer by using a cloth dampened with a suitable disinfectant such as 10% bleach or isopropyl alcohol. Be sure that no liquid enters the analyzer.
- Clean the test strip tray every day.

### Procedure

1. Turn off the power.
2. Remove the test strip tray.
3. Rinse the tray under running water.
4. Remove any crystalline deposits, especially those contaminating the retaining bar mechanism, with a soft brush.
5. Wipe down the test strip tray with isopropyl alcohol or 10% bleach.
6. Dry with a soft, lint-free cloth.
7. Install the test strip tray by holding the tray opposite the side with the gray reference pad and inserting the tray into the slot below the function keys. Do not touch the gray reference pad.
8. Switch on the instrument. During the self-check process, the system will verify that the reference pad is in suitable condition for reading test strips. If not, an error message will be displayed.

### Policy – Replacing Printer Paper

Replace the printer paper **the day it runs out**, following the instructions in Section 4.2 of the *Urisys 1100<sup>®</sup> Operator's Manual*. This will ensure a paper print-out of calibration results will be available the next day if calibration is performed upon instrument start up.

## Policy – Documentation and Replacement

- Any maintenance, calibration or replacement is documented on the *Preventive Maintenance Log*.
- Replaceable parts are available. They can be ordered from your local Roche Diagnostics distributor or are available from \_\_\_\_\_ (location in your facility).

Replaceable parts for the Urisys 1100® Analyzer are:

Replaceable Part	Catalog Number
Test strip tray - Type C for Chemstrip 10MD	3666735
Test strip tray - Type N for Chemstrip 5 OB/Chemstrip 7	3666913
AC Adapter	4340647
Power cord	4340612
Reference pads	1907131
Thermal paper	3666751
Interface connection cable	1906186
Keyboard	1248693
Barcode reader w/o trigger	???
Barcode reader cable	???

#See your Operator's Manual to determine which test strip tray is in your analyzer.

- If staff are unable to correct a problem with the Urisys 1100® Urine Analyzer, they are instructed to contact \_\_\_\_\_ (location in your facility) or call Roche Diagnostics Technical Service Center at 1-800-428-4674.
- The appointed individual reviews the *Preventive Maintenance Log* for completion \_\_\_\_\_ (when? weekly, monthly, quarterly).
- Insert your policy for who is responsible for performing and documenting preventive maintenance.

Name

Title

---



---



---



---



---



---



---



---



---



---

### III. Infection Control Guidelines

#### Policy

- Because of the potentially hazardous nature of handling urine and the possibility of its contamination with blood, it is recommended that disposable latex gloves and laboratory coats be used when collecting specimens and performing test procedures. Consult your facility's infection control policy.
- Disposable latex gloves and laboratory coats are to be removed and hands washed thoroughly with soap and water after completing the test procedure and prior to handling equipment not related to the procedure.
- Urine specimens should be handled at Biosafety Level 2 as recommended for any potentially infectious material in the Centers for Disease Control/National Institutes of Health manual, *Biosafety in Microbiological and Biomedical Laboratories*, 1988. Universal precautions may apply if the urine is contaminated with blood or if required by the infection control policy of your facility.<sup>24</sup>
- If a spill should occur, disinfect the area with a 1:10 bleach solution (1 part bleach plus 9 parts water) or follow the infection control procedures of your facility.

## IV. Operator Certification/Recertification

### Policy

- Operators receive appropriate certification to perform tests with the Urisys 1100® Urine Analyzer according to your facility's policy.
- Certification is documented and maintained on the *Operator Certification Log* and *Skills Checklist*. A *Knowledge Test* is also provided.
- Only certified operators may use the Urisys 1100® Urine Analyzer.
- Operator competency checks (recertification) will be performed *yearly* and recorded on the *Operator Certification Log* and *Operator Competency Log*.

### Procedure – Operator Competency

The following procedure may be used to certify/recertify your operators and check their competency on the use of the Urisys 1100® Urine Analyzer.

*Note: A valid calibration is required before performing the following within-run precision.*

#### Within-Run Precision

1. The following equipment is assembled prior to testing:

Urisys 1100® Urine Analyzer

Chemstrip urine test strips

Quality control material–1, 2, or 3 levels (according to your facility's policy) prepared according to the manufacturer's instructions

Disposable latex gloves (if required by facility)

Laboratory coat (if required by facility)

Absorbent paper

*Operator Competency Log*

2. Run control 20 times consecutively on the Urisys 1100® Urine Analyzer. Refer to *Section I.H. Quality Control Procedure* for instructions.
3. Record results on the *Operator Competency Log*.
4. Determine the result that occurs most frequently (mode) for each parameter and record the result.

5. Determine the color block variation (plus or minus the number of color blocks from the mode) for each parameter in the run of 20. Record the result for each parameter.
6. The color block variation (number of color blocks away from the mode) for each chemistry of the 20 analyses should be no greater than  $\pm$  one color block. Refer to *Section 1.H. Quality Control Procedure* if the color block variation exceeds plus or minus one.

Example 1: If 20 consecutive glucose results produce 20 results of 100 mg/dL, the mode is 100 mg/dL and the color block variation is 0. This is acceptable performance.

Example 2: If 20 consecutive glucose results produce 15 results of 100 mg/dL and 5 results of 250 mg/dL, the mode is 100 mg/dL and the color block variation is + one color block. This is acceptable performance.

Example 3: If 20 consecutive glucose results produce two results of 50 mg/dL, 16 results of 100 mg/dL and two results of 250 mg/dL, the mode is 100 mg/dL and the color block variation is + one and - one color block. This is acceptable performance.

Example 4: If 20 consecutive glucose results produce two results of “normal”, 16 results of 100 mg/dL and two results of 250 mg/dL, the mode is 100 mg/dL and the color block variation would be + one and - two color blocks. This is unacceptable performance.

## Control Recovery

1. Record the control package insert acceptable ranges for the Urisys 1100<sup>®</sup> Urine Analyzer on the *Operator Competency Log*. The Urisys 1100<sup>®</sup> results should be compared to the *visual* acceptable ranges if the Urisys 1100<sup>®</sup> Analyzer is not referenced on your control package insert.
2. Compare the acceptable ranges stated in the package insert to the results obtained from the within-run precision step 4.
3. The mode results should fall within the acceptable control ranges. Refer to *Section 1.H. Quality Control Procedure* for instructions if the results fall outside the acceptable control ranges.

## Guidelines

Insert your policy for who is responsible for training and documenting operator certification / recertification with the Urisys 1100<sup>®</sup> Urine Analyzer.

---

---

---

## References

1. Kolthoff, I. *Acid-Base Indicators*. McMillan Co., New York, 1937, p. 171.
2. Free, H., G. Collins and A. Free. Triple-test Strip for Urinary Glucose, Protein and pH, *Clin Chem*, 1960, 6:352.
3. Griess, P., Notes on the Paper of Weselsky and Benedikt. Some Azo Compounds. *Ber. Dtsch. Chem. Ges.*, 1879, 12:426.
4. Weltmann, O., Method for the Simple Detection of Urinary Tract Infections. *Wien. Med. Wschr*, 1922, 72:618.
5. Fuchs, T. and G. Gutensohn. Value and Limitation of the Nitrite Test in Diagnosis of Pyelonephritis. *Dtsch. Med. J.*, 1967, 10:343.
6. Sørensen, S. The Measurement of the Hydrogen Ion Concentration and Its Importance for Enzymatic Processes. *Biochem. Z.*, 1909, 21:131.
7. US Patent No. 4013416.
8. Gyure, W. L. Comparison of Several Methods for Semiquantitative Determination of Urinary Protein. *Clin. Chem*, 1977, 23:876.
9. Keston, A. Abstracts of papers presented at the 129th meeting of the American Chemical Society, Dallas, April, 1956, p. 31c.
10. Comer, J. Semiquantitative Specific Test Paper for Glucose in Urine. *Anal. Chem.*, 1956, 28:1748.
11. Legal, E. A New Acetone Reaction and its Applicability for the Examination of Urine. *Chem. Centr.*, 1883, 14:652.
12. Chertack, M. and J. Sherrick. Evaluation of A Nitroprusside Dip Test For Ketone Bodies. *J. A. M. A.*, 1958, 167:1621.
13. Kutter, D. et al. Usefulness of a New Test Strip for Detecting Urobilinogen in Urine. *Dtsch. Med. Wschr.*, 1973, 98:112.
14. With, T. *Bile Pigments, Chemical, Biological and Clinical Aspects*. Academic Press, New York, 1968, p. 492.
15. Cook, M., et al. The Detection of Blood in Urine. *Am. J. Med. Tech.*, 1956, 22:218.
16. Leonards, J. Simple Test For Hematuria Compared With Established Tests. *J. A. M. A.*, 1962, 179:807.
17. Braun, J. and W. Straube. A New Rapid Test For Diagnosing Microhematuria, Compared With Results Of Microscopic Examination. *Dtsch. Med. Wschr.*, 1975, 100:87.
18. Kutter, D. et al. Usefulness of a New Test Strip for the Detection of Erythrocytes and Hemoglobin in Urine. *Dtsch. Med. Wschr.*, 1974, 99:2332.
19. Brühl, P. et al. Clinical Experience Gained with a Test Strip Used to Screen for Erythrocytes and Hemoglobinuria. *Therapiewoche*, 1976, 26:5193.
20. NCCLS. Protection of Laboratory Workers from Infectious Disease Transmitted by Blood and Tissue. *NCCLS Doc. M29-P*, 7, 9, May, 1985, 342-347.

21. Csako, G. False Positive Results for Ketone with the Drug Mesna and Other Free-Sulfhydryl Compounds. *Clin. Chem.*, 1987, 33/2:289.
22. Henry, J.B. et al. *Clinical Diagnosis and Management by Laboratory Methods*, 17th ed. Saunders Philadelphia, 1984, 394;1441.
23. Czerwinski, A. W., et al. Further Evaluation of the Griess Test to Detect Significant Bacteriuria. *Amer. J. Obstet. Gynec.*, 1971, 110:677.
24. Centers for Disease Control/ National Institutes of Health. *Biosafety in Microbiological and Biomedical Laboratories*, 1988.

## **V. Appendix**

### **Instructions for Collections of a Midstream Urine**

When examining urine specimens, every precaution must be taken to avoid contamination. Therefore, make sure your patients follow the instructions below carefully so that reasonable test results are obtained.

It is important to note that some cleansing towelettes may use soaps that contain oxidizing agents. Oxidizing agents may interfere with some parameters on the urine test strips. Therefore, it is extremely important that thorough rinsing take place after cleansing the genitals.

### **Procedure**

1. Do not open the specimen container until immediately prior to use.

2. Procedure

#### **Female**

- Kneel or squat over a bedpan or stand astride a toilet bowl.
- Separate labia minor.
- Cleanse meatus (opening of urethra) three times.
- Rinse area with sterile, water-saturated sponge. Do not dry.
- Void forcibly, allowing initial stream of urine to pass into the lavatory.
- Collect midstream in sterile container.

#### **Male**

- Cleanse glans three times.
- Rinse area with sterile, water-saturated sponge. Do not dry.
- Void forcibly, allowing initial stream of urine to pass into the lavatory.
- Collect midstream in sterile container.

## **Infant**

- Cleanse genitalia three times.
  - Rinse area with sterile, water-saturated sponge. Do not dry.
  - Supporting infant in a face down position, stimulate urination by stroking paraspinal muscles.
  - If infant fails to void, place sterile bag over genitals.
  - If infant does not void within 45 minutes, re-prepare the area and apply a new bag.
  - Collect urine in a sterile container.
  - Do not allow the container to touch the body.
3. Close the container and present it immediately for urinalysis.

## Routine Urinalysis Controls Compatible with the Urisys 1100® Urine Analyzer

Several commercial controls are available. Controls may vary in number of levels or components, necessity for reconstitution or ready to use, type and volume of container, and usefulness as a control for microscopic examination.

When selecting a control, consider the above-mentioned variables to identify the control most suitable for your facility's needs. Controls typically can be ordered through medical/surgical supply distributors or directly from the manufacturer.

Refer to the following table to determine which control best meets the needs of your facility's quality control requirements.

Urine Controls Chart				
Brand Name	KOVA-Trol® Urinalysis Control	KOVA® Liqua-Trol™ Urinalysis Control	The Dipper Urine Dipstick Control	Liquichek™ Routine Urine Chemistry Control
Manufacturer	HYCOR Biomedical Inc.	HYCOR Biomedical Inc.	Quantimetrix Corporation	BIO RAD
Website	<a href="http://www.hycorbiomedical.com">www.hycorbiomedical.com</a>	<a href="http://www.hycorbiomedical.com">www.hycorbiomedical.com</a>	<a href="http://www.4qc.com">www.4qc.com</a>	<a href="http://www.bio-rad.com">www.bio-rad.com</a>
Address	7272 Chapman Ave. Garden Grove, CA 92641	7272 Chapman Ave. Garden Grove, CA 92641	2005 Manhattan Beach Bou- levard Redondo Beach, CA 90278	4000 Alfred Nobel Dr Hercules, CA 94547
Toll-free No.	800-382-2527	800-382-2527	800-624-8380	800-224-6723
Phone No.	714-895-9558	714-895-9558	310-536-0006	510-724-7000
Fax No.	714-933-3222	714-895-6920	800.845.1834	510-741-6373
Product Information Requests	<a href="http://www.hycorbiomedical.com/site/inforequest.html">http://www.hycorbiomedical.com/site/inforequest.html</a>	<a href="http://www.hycorbiomedical.com/site/inforequest.html">http://www.hycorbiomedical.com/site/inforequest.html</a>	<a href="mailto:info@4qc.com">info@4qc.com</a>	<a href="mailto:diagcs@bio-rad.com">diagcs@bio-rad.com</a>
Distribution	Cardinal, Fisher, PSS, McKesson, Owens & Minor, Henry Schein	Cardinal, Fisher, PSS, McKesson, Owens & Minor, Henry Schein	Direct, online thru website , Cardinal Health, Fisher, Infolab, Labsco, McKesson, Owens & Minor, PSS	Direct, online catalog on website, Cardinal, PSS, Labsco
Number of Levels	3: Negative Norm Low Abnormal High Abnormal/ High Abnormal w/urob.	2: Normal Abnormal	2: Normal Abnormal	2: Level 1- Negative/Norm. Level 2- Positive/Abnormal
Preparation	Reconstitution	Ready to use	Ready to use	Ready to use
Package Size	4 x 15 mL 4 x 60 mL 10 x 60 mL	3 x 15 mL	6 x 15 mL	10 x 10 mL
Stability / Unopened	Until exp. date on vial (18mo.)	Until exp. date on vial	Until exp. date on vial	Until exp. date on vial
Stability / Opened	5-7 days after recon 4 mo. frozen	Until exp. date on vial	90 days 2-8°C	7 days 2-8°C
Microscopic Values	Yes	Yes	No	No

## VI. Log Sheets

This section contains the following blank log sheets that may be copied and used for documentation for the Urisys 1100<sup>®</sup> Urine Analyzer in your facility.

*Operator Certification Log* – for recording the names of trained Urisys 1100<sup>®</sup> operators and tracking yearly competency checks.

*Operator Competency Log* – for recording yearly competency checks for each operator, within-run precision data, and control recovery data.

*Reagent Log* – for recording the date received, lot numbers and expiration dates of Chemstrip urine test strips, Chemstrip Calibration Strips and controls.

*Urinalysis Report Form* – for recording patient results.

*Quality Control Log* – for recording quality control information.

*Preventive Maintenance Log* – for recording the dates on which maintenance was performed.

*Temperature Log (2)* – for recording Urisys 1100<sup>®</sup> Urine Analyzer operating temperatures and Chemstrip urine test strip storage temperatures.

*Knowledge Test* – for written documentation of operator certification.

*Skills Checklist* – for documenting operator skills while training on the Urisys 1100<sup>®</sup> Urine Analyzer



# Operator Competency Log

## Urisys 1100<sup>®</sup> Urine Analyzer

Operator Name: \_\_\_\_\_ Date: \_\_\_\_\_

Chemstrip Cat. No. \_\_\_\_\_ Lot No.: \_\_\_\_\_ Exp. Date: \_\_\_\_\_

Control Name and Level: \_\_\_\_\_ Lot No.: \_\_\_\_\_ Exp. Date: \_\_\_\_\_

### Within-Run Precision

	SG	pH	LEU	NIT	PRO	GLU	KET	UBG	BILI	BLD
1										
2										
3										
4										
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										
16										
17										
18										
19										
20										
Mode										
Color block variation										

Note: The color block variation should be no greater than +/- one color block from the mode for acceptable performance.

# Operator Competency Log

## Urisys 1100® Urine Analyzer

### Control Recovery

	Control Package Insert Ranges	Urisys 1100® Urine Analyzer Mode Result from Within Run Precision
SG		
pH		
LEU		
NIT		
PRO		
GLU		
KET		
UBG		
BILI		
BLD		

Note: Reference the visual acceptance range if the Urisys 1100 Urine Analyzer is not referenced on your control package insert.



# Urinalysis Report Form

## Urisys 1100<sup>®</sup> Urine Analyzer

Patient Name: \_\_\_\_\_

Age: \_\_\_\_\_ M F

Physician's Name: \_\_\_\_\_

Collection Date: \_\_\_\_\_ Test Date: \_\_\_\_\_ Operator's Initials: \_\_\_\_\_

### Physical Examination

Color:	colorless	yellow	amber	other
Appearance:	clear	hazy	cloudy	turbid

### Chemical Examination

Specific Gravity	1.000	1.005	1.010	1.015	1.020	1.025	1.030
pH	5	6	6.5	7	8	9	
Leukocytes	neg	25	75	500			
Nitrite	neg	pos	(any pink color is considered positive)				
Protein	neg	trace	30	100	500		
Glucose	norm	50	100	250	>1000		
Ketones	neg	15	50	150			
Urobilinogen	norm	1	4	8	>12		
Bilirubin	neg	1	3	6			
Blood	neg	trace	50	250			

Comments:

\_\_\_\_\_

\_\_\_\_\_

Attach the Patient Report printout here if required by your facility.

# Quality Control Log

## Urisys 1100<sup>®</sup> Urine Analyzer

Name of control solution used by your facility: \_\_\_\_\_

Control Lot No.: \_\_\_\_\_ Exp. Date: \_\_\_\_\_

Chemstrip Cat. No.: \_\_\_\_\_

Chemstrip Lot No.: \_\_\_\_\_ Exp. Date: \_\_\_\_\_

Operator's Initials: \_\_\_\_\_

	Date	SG	pH	LEU	NIT	PRO	GLU	KET	UBG	BILI	BLD	Yes/ No	Initials	Corrective Action
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														
Level 1														
Level 2														

Note: Reference the visual acceptable range if the Urisys 1100<sup>®</sup> Urine Analyzer is not referenced on your control package insert.



# Temperature Log

## Urisys 1100<sup>®</sup> Urine Analyzer

Location: \_\_\_\_\_

Year: \_\_\_\_\_

Acceptable Range (°C): \_\_\_\_\_

	Jan	Feb	Mar	Apr	May	June	July	Aug	Sept	Oct	Nov	Dec
1												
2												
3												
4												
5												
6												
7												
8												
9												
10												
11												
12												
13												
14												
15												
16												
17												
18												
19												
20												
21												
22												
23												
24												
25												
26												
27												
28												
29												
30												
31												

# Temperature Log

## Chemstrip Urine Test Strips

Cat. No. \_\_\_\_\_

Location: \_\_\_\_\_

Year: \_\_\_\_\_

Acceptable Range (°C): \_\_\_\_\_

	Jan	Feb	Mar	Apr	May	June	July	Aug	Sept	Oct	Nov	Dec
1												
2												
3												
4												
5												
6												
7												
8												
9												
10												
11												
12												
13												
14												
15												
16												
17												
18												
19												
20												
21												
22												
23												
24												
25												
26												
27												
28												
29												
30												
31												

# Knowledge Test

## Urisys 1100® Urine Analyzer

Name: \_\_\_\_\_ ID No.: \_\_\_\_\_

Date: \_\_\_\_\_ Unit: \_\_\_\_\_

Mark "T" if the statement is true and "F" if the statement is false.

- \_\_\_\_\_ 1. You must check the expiration dating on the vial each time you perform a patient or quality control test.
- \_\_\_\_\_ 2. Dip the Chemstrip urine test strip briefly (no longer than 1 second) in the urine sample.
- \_\_\_\_\_ 3. Strong oxidizing cleaning agents interfere with the blood patch results.
- \_\_\_\_\_ 4. Correct placement of the Chemstrip urine test strip in the tray of the analyzer is necessary to obtain accurate results.
- \_\_\_\_\_ 5. It is necessary to bring refrigerated samples to room temperature before testing.
- \_\_\_\_\_ 6. The sample should be in sufficient volume to allow all test pads to be completely wetted.
- \_\_\_\_\_ 7. Running the side of the Chemstrip urine test strip along the rim of the container removes excess sample.
- \_\_\_\_\_ 8. If you do not obtain a successful calibration, the analyzer will still process samples.
- \_\_\_\_\_ 9. A positive nitrite test is not normal in fresh, clean-catch midstream urine specimens.
- \_\_\_\_\_ 10. Urine residue on the tray of the analyzer may result in inaccurate results.
- \_\_\_\_\_ 11. Green spots on the blood patch are indicative of unlysed erythrocytes.

**Fill in the blanks.**

12. The specimen should be tested within \_\_\_\_\_ hour(s) of collection, otherwise, refrigerate it in a closed container.
13. Replace the \_\_\_\_\_ to avoid deterioration of the test strips.
14. When using Chemstrip 10MD Test Strips, the Urisys 1100® Urine Analyzer must be calibrated every \_\_\_\_\_ days.
15. Reasonable amounts of ascorbic acid do not affect the \_\_\_\_\_ test.
16. The specimen must be thoroughly \_\_\_\_\_, as red cells sink to the bottom and could cause false positive results.
17. Rinse the test strip tray with a suitable disinfectant \_\_\_\_\_ (when?).

## Answer Key

1. T
2. T
3. T
4. T
5. T
6. T
7. T
8. F
9. T
10. T
11. T
12. one
13. vial cap
14. seven
15. blood or glucose
16. mixed
17. daily

# Skills Checklist

## Urisys 1100™ Urine Analyzer

The trainer should check each activity as it is demonstrated. All activities must be completed before the trainee is certified to use the Urisys 1100® Urine Analyzer.

### Specimen Collection

1. User assembles specimen collection supplies
  - Specimen container
  - Cleansing towelette
  - Water-saturated sponge
2. User reviews specimen collection
  - States patient preparation steps
  - Explains specimen storage and stability
  - Describes clean-catch, midstream urine technique

### Testing Procedure

1. User assembles testing supplies
  - Urisys 1100® Urine Analyzer
  - Chemstrip Urine Test Strips
  - Chemstrip Urine Analyzer Calibration Strips
  - Patient specimen
  - Disposable latex gloves and laboratory coat
  - Absorbent paper
2. User preparation
  - Checks expiration date on test strip vial
  - Puts on disposable latex gloves and laboratory coat

### 3. Sample preparation

- States if refrigerated, bring sample to room temperature
- Mixes sample by swirling capped container several times

### 4. Calibration (*when using Chemstrip 10MD or Chemstrip 10UA Test Strips*)

- Removes one Chemstrip urine analyzer calibration strip
- Does not touch elevated gray areas on strip
- Places calibration strip(pad side up) onto the test strip tray
- Presses START button
- Documents calibration on the *Preventive Maintenance Log*

### 5. Testing

- Removes one Chemstrip urine test strip
- Replaces vial cap immediately
- Ensures analyzer test strip tray is clean
- Briefly dips test strip vertically into specimen
- Draws strip edge along rim of container to remove excess urine
- Touches strip to absorbent paper
- Places the test strip (pad side up) onto the test strip tray
- The handle end of the strip should be held by the clip at the insertion slot
- Obtains results from the printout
- Removes test strip
- Wipes tray with dry lint-free cloth. If there is any residue on the tray, follows proper maintenance as outlined in the *Urisys 1100® Urine Analyzer Operator's Manual*.

6. Results

- Documents results as appropriate for your facility

7. Quality control

- States when to perform QC tests
- Performs test like routine urine.
- Documents results
- Reviews procedure if the test results fall outside the acceptable range
- If problem persists, calls Roche Diagnostics Technical Service Center at 1-800-428-4674

8. Infection control

- Properly disposes of specimen, used test strips and container
- Cleans and decontaminates the work area according to infection control policy
- Removes and disposes of disposable latex gloves

---

Operator Signature

Date

---

Trainer Signature

Date

**Contact your Roche Diagnostics Account Manager for training certificates.**

# VII. Regulatory Requirements

# VIII. Product Information

CE

 Roche Diagnostics Corporation  
9115 Hague Road  
Indianapolis, IN 46256

[www.poc.roche.com](http://www.poc.roche.com)  
[www.diafant.com](http://www.diafant.com)

Cat. No. 3755541  
REF 0 3755541001

